


# Current Sensor CLV30

Model CLV30-500A-3.5VAC-60Hz

Serial 990428-1

External load of 345 at end of 25' T.S.P.

Data taken on 04/28/99



Amps	volts	fit	error	phase
0.00	0.00	0.00	0.0%	0.7
49.68	0.47	0.35	3.6%	1.2
99.36	0.90	0.70	5.7%	1.7
198.71	1.65	1.39	7.4%	2.0
298.07	2.31	2.09	6.2%	2.0
397.43	2.91	2.78	3.7%	2.0
496.78	3.50	3.48	0.6%	2.2
596.14	4.00	4.17	-4.9%	2.5

**Smith Research & Technology, Inc.**



3109 N. Cascade Ave., Bldg. #201  
 Colorado Springs, CO 80907-5190  
 719 634 2259, FAX 719 634 2601