


Current Sensor CQ25

Model CQ25-50A-3.5VAC-60Hz

Serial 931211-1

2.24K ext.load

Data taken on 11 Dec. 1993



Amps	volts	fit	error	phase
0.00	0.00	0.00	0.0%	7.9
4.98	0.37	0.35	0.5%	5.7
9.95	0.75	0.70	1.6%	3.5
19.90	1.49	1.39	2.7%	6.4
29.85	2.18	2.09	2.6%	9.2
39.82	2.85	2.79	1.7%	11.1
49.75	3.50	3.48	0.4%	12.4
59.70	4.13	4.18	-1.4%	13.6

Smith Research & Technology, Inc.



3109 N. Cascade Ave., Bldg. #201
 Colorado Springs, CO 80907-5190
 719 634 2259, FAX 719 634 2601