


Current sensor, DC out, CQ32

CQ32-10A-5VDC-60Hz
 Serial 970623-2
 Data taken on 06/24/97

RDC=60k , Ripple=24mV p-p



amps	volts	fitted	error
0.00	0.00	0.00	0.00%
1.00	0.50	0.50	-0.04%
2.00	1.00	1.00	0.00%
4.00	2.00	2.00	0.00%
6.00	3.00	3.00	-0.01%
8.00	3.99	4.00	-0.12%
10.00	5.00	5.00	0.04%
12.00	6.02	6.00	0.42%

Smith Research & Technology, Inc.



3109 N. Cascade Ave., Bldg. #201
 Colorado Springs, CO 80907-5190
 719 634 2259, FAX 719 634 2601