



AC-output voltage sensor

Model VDQ277MB-3.5VAC-60Hz
 Serial 931011-6
 Data taken on 10/27/93




Input volts	Output volts	fit	error	phase
221.40	2.85	2.86	0.00	9.60
249.30	3.18	3.18	0.00	10.20
276.90	3.50	3.50	0.00	10.90
304.50	3.81	3.81	0.00	11.60
332.30	4.12	4.13	0.00	12.40
156.00	4.45	4.48	0.00	8.20

Smith Research & Technolog



3109 N. Cascade Ave., Bldg
 Colorado Springs, CO 809
 719 634 2259, FAX 719 6



ogy, Inc.
Bldg. #201
0907-5190
) 634 2601