



AC-output voltage sensor

Model VDQ480MB-3.5VAC-60Hz
 Serial 980409-3
 Data taken on 04/13/98




Input volts	Output volts	fit	error	phase
384.00	2.85	2.86	0.00	10.20
432.00	3.18	3.18	0.00	11.20
480.00	3.50	3.50	0.00	11.90
528.00	3.82	3.82	0.00	12.90
576.00	4.12	4.14	0.00	14.20

Smith Research & Technology



3109 N. Cascade Ave., BI
 Colorado Springs, CO 80909
 719 634 2259, FAX 719 634 2258



ogy, Inc.

Bldg. #201
30907-5190
9 634 2601